

From: [Moore, Gary](#)
To: [Gregg, Diane](#)
Subject: Re: Assistance Needed
Date: Wednesday, November 05, 2014 8:26:00 AM

Christy:

At the present, I am really looking at FT1004 and FT506. I may need additional help on others. If I can bring in a sample would you guys be able to do this for me?

Thanks

Gary Moore

Federal On-Scene Coordinator

U.S. EPA Region 6

214-789-1627 cell

214-665-6609 office

moore.gary@epa.gov

From: Gregg, Diane
Sent: Wednesday, November 5, 2014 8:18 AM
To: Warren, Christy
Cc: Moore, Gary; Flores, Raymond
Subject: RE: Assistance Needed

I'm not sure I completely understand the table; however, it appears that most of the containers with codes that could not be excluded have sludge or oil in the consistency description. I am not surprised that the presence of oil would cause issues for 2,4-DNT and Hexachlorobenzene since these compounds have the lowest MCLs under TCLP. There's only so much oil you can put through a GC/MS before it makes the analysis unusable. Also, 2,4-DNT has quant and qual masses that are usually present in oil (in that typical hydrocarbon hump) so analyzing this by SIM probably wouldn't yield any better data. It's too difficult to distinguish from background in the oil. HCB does have unique masses and SIM should help. As for Chlordane, if the lab performs a sulfuric acid cleanup on the sample extract, they should remove most of the interferences from the oil and the analysis should be relatively easy after that cleanup. The only option I can think for 2,4-DNT is GC/QQQ. This analysis creates daughter ions that may be unique enough to quantify. Just keep in mind we did this for Norphlet (aka MacMillan Ring free) and the masses for some PAHs were not unique enough from the oil. This is for DNT so it might work. We'd never know until we tried.

From: Warren, Christy
Sent: Wednesday, November 05, 2014 7:59 AM
To: Gregg, Diane
Cc: Moore, Gary; Flores, Raymond
Subject: FW: Assistance Needed

Diane,

Can you help Gary?

2,4 Dinitrotoluene is D030, Hexachlorobenzene is D032 and Chlordane is D020

Christy Warren

U.S. EPA Region 6 Laboratory

Sample Control Coordinator

10625 Fallstone Road

Houston, Texas 77099

281-983-2137 (office)

281-983-2124 (fax)

281-415-6815 (cell)

713-805-7157 (cell)

Confidentiality Notice: This e-mail, including all attachments, is for the sole use of the intended recipient(s) and may contain confidential and privileged information. Any unauthorized review, use, disclosure or distribution is prohibited. If you are not the intended recipient, please contact the sender and destroy all copies of this message

From: Moore, Gary

Sent: Tuesday, November 04, 2014 7:05 PM

To: Flores, Raymond; Warren, Christy

Subject: CES: Assistance Needed

Ray/Christy:

I need some lab assistance for this site that I am working in Houston. I have having difficulty getting the commerical lab to meet the TCLP regulatory levels due to matrix interferences. I am sending you this database so that you can see where we are having difficulty. It is primarily for 2,4-Dinitrotoluene, Hexachlorobenzene, Chlorodane. Some that we have talked to have talked about SIM Analyses. Do you know what we can do to resolve these issues?

See Table

Thanks

Gary Moore

Federal On-Scene Coordinator

U.S. EPA Region 6

214-789-1627 cell

214-665-6609 office

moore.gary@epa.gov